

## Plasma/Serum Circulating RNA Purification Maxi Kit (Slurry Format)

Product # 50900

Product Insert

Norgen's Plasma/Serum Circulating RNA Purification Maxi Kit (Slurry Format) provides a fast, reliable and simple procedure for isolating circulating RNA from various amounts of plasma/serum ranging from 2 mL to 5 mL. Free-circulating RNA in plasma and serum has the potential to provide biomarkers for certain cancers and disease states, and includes tumor-specific extracellular RNA in the blood. Norgen's Plasma/Serum Circulating RNA Purification Maxi Kit (Slurry Format) provides an efficient method for the purification of all sizes of these fragmented free-circulating RNAs from human plasma or serum.

Purification is based on the use of Norgen's proprietary resin as the separation matrix. The kit is able to isolate all sizes of circulating RNA, including microRNA. Norgen's Plasma/Serum Circulating RNA Purification Kit (Slurry Format) provides an advantage over other available kits in that it does not require extension tubes for the purification of free-circulating RNA from large sample volumes. RNA can be isolated from either fresh or frozen samples using this kit, and the kit allows for the concentration of RNA that is present in low concentrations (1-100ng/mL circulating RNA in human plasma). Typical yields of free-circulating RNA will vary depending on the input sample, as the amount of RNA present in plasma and serum will depend upon the health status of the individual and the level of nucleases present in the blood.

***This kit is suitable for the isolation of RNA from serum or plasma prepared from blood collected on either EDTA or citrate. Plasma samples prepared from blood collected on heparin should not be used as heparin can significantly interfere with many downstream applications such as RT-PCR.***

Preparation time for a 10 samples is less than 40 minutes. The purified plasma/serum free-circulating RNA are eluted in an elution solution that is compatible with reverse transcription qPCR, reverse transcription PCR, Northern blotting, RNase protection and primer extension, and expression array assays.

### Kit Components:

| Component                | Contents   |
|--------------------------|------------|
| PS Solution A            | 6 mL       |
| PS Solution B            | 2 x 125 mL |
| PS Solution C            | 9 mL       |
| Wash Solution            | 11 mL      |
| RNA Elution Solution     | 3 mL       |
| Mini Filter Spin Columns | 25         |
| Collection Tubes         | 25         |
| Elution tubes (1.7 mL)   | 25         |
| Product Insert           | 1          |

### Customer-Supplied Reagents and Equipment

- Centrifuge with a swinging bucket rotor capable of 2000 RPM
- Benchtop microcentrifuge
- Micropipettors
- 96 – 100% ethanol
- β - Mercaptoethanol
- 50 mL tubes
- 15 mL tubes

### Storage Conditions and Product Stability

All buffers should be kept tightly sealed and stored at room temperature (15-25°C) for up to 1 year without showing any reduction in performance.

It is recommended to warm PS Solution A, PS Solution B and PS Solution C for 20 minutes at 60°C if any salt precipitation is observed.

### Quality Control

In accordance with Norgen's Quality Management System, each lot of Norgen's Plasma/Serum Circulating RNA Purification Maxi Kit (Slurry Format) is tested against predetermined specifications to ensure consistent product quality.

### Product Use Limitations

Norgen's Plasma/Serum Circulating RNA Purification Maxi Kit (Slurry Format) is designed for research purposes only. It is not intended for human or diagnostic use.

### Product Warranty and Satisfaction Guarantee

NORGEN BIOTEK CORPORATION guarantees the performance of all products in the manner described in our product manual. The customer must determine the suitability of the product for its particular use.

### Safety Information

Ensure that a suitable lab coat, disposable gloves and protective goggles are worn when working with chemicals. For more information, please consult the appropriate Material Safety Data Sheets (MSDSs). These are available as convenient PDF files online at [www.norgenbiotek.com](http://www.norgenbiotek.com).

**CAUTION: DO NOT add bleach or acidic solutions directly to the sample-preparation waste.**

**PS Solution A, PS Solution B** and **PS Solution C** contain guanidinium salts, and should be handled with care. Guanidinium salts form highly reactive compounds when combined with bleach, thus care must be taken to properly dispose of any of these solutions

Plasma or serum of all human and animal subjects is considered potentially infectious. All necessary precautions recommended by the appropriate authorities in the country of use should be taken when working with plasma or serum.

### Working with RNA

RNases are very stable and robust enzymes that degrade RNA. Autoclaving solutions and glassware is not always sufficient to actively remove these enzymes. The first step when preparing to work with RNA is to create an RNase-free environment. The following precautions are recommended as your best defence against these enzymes.

- The RNA area should be located away from microbiological work stations
- Clean, disposable gloves should be worn at all times when handling reagents, samples, pipettes, disposable tubes, etc. It is recommended that gloves are changed frequently to avoid contamination
- There should be designated solutions, tips, tubes, lab coats, pipettes, etc. for RNA only
- All RNA solutions should be prepared using at least 0.05% DEPC-treated autoclaved water or molecular biology grade nuclease-free water
- Clean all surfaces with commercially available RNase decontamination solutions
- When working with purified RNA samples, ensure that they remain on ice during downstream applications

## Purification of Circulating RNA from 2mL Serum or Plasma

### Notes Prior to Use

- All centrifugation steps are performed at room temperature.
- A variable speed centrifuge should be used for maximum kit performance. If a variable speed centrifuge is not available a fixed speed centrifuge can be used, however reduced yields may be observed.
- Ensure that all solutions are at room temperature prior to use.
- Prepare a working concentration of the **Wash Solution** by adding 25 mL of 96-100% ethanol (provided by the user) to the supplied bottle containing the concentrated RNA Wash Solution. This will give a final volume of 36 mL. The label on the bottle has a box that may be checked to indicate that the ethanol has been added.
- The use of  $\beta$ -mercaptoethanol in lysis is highly recommended to isolate RNA for sensitive downstream applications. Add 10  $\mu$ L of  $\beta$ -mercaptoethanol (provided by the user) to each 1 mL of PS Solution B..

- **PS Solution A contains resin and must be mixed well before every pipetting.**
  - It is highly recommended to warm up **PS Solution A**, **PS Solution B** and **PS Solution C** at 60°C for 20 minutes and mix well until the solutions become clear again if precipitates are present.
  - It is important to work quickly during this procedure.
  - **This kit is suitable for the isolation of RNA from serum or plasma prepared from blood collected on either EDTA or citrate. Plasma samples prepared from blood collected on heparin should not be used as heparin can significantly interfere with many downstream applications such as RT-PCR.**
  - **The procedure is outlined for 2 mL inputs. To process different Plasma/Serum volumes please check Table 1 for the appropriate volumes that should be added from PS Solution A, PS Solution B and 96-100% Ethanol to different Plasma/Serum sample volumes. The volume of PS Solution C is fixed for all Plasma/Serum volumes.**
1. In a 50 mL tube (provided by the user), add 0.2 mL of **PS Solution A** and 3.8 mL **PS Solution B** (after the addition of  $\beta$ -mercaptoethanol) for every 2 mL plasma/serum sample. Mix well by vortexing for 15 seconds. (**Note: PS Solution A contains resin and must be mixed well before every pipetting**)
  2. Incubate the mixture from **Step 1** for 10 minutes at 60°C.
  3. After incubation add 6 mL of 96-100% Ethanol (provided by the user). Mix well by vortexing for 15 seconds.
  4. Centrifuge for **30 seconds at 1,000 RPM**, then carefully decant the supernatant in order to ensure that the slurry pellet is not dislodged.
  5. To the slurry pellet add 0.3 mL **PS Solution C**, and mix well by vortexing for 15 seconds
  6. Incubate the mixture from Step 5 for 10 minutes at 60°C.
  7. After incubation add 0.3 mL 96-100% Ethanol (provided by the user). Mix well by vortexing for 15 seconds.
  8. Transfer 650  $\mu$ L from the mixture from **Step 7** into a Mini Filter Spin column. Centrifuge for **1 minute at 14,000 RPM**. Discard the flowthrough and reassemble the spin column with its collection tube;
  9. Repeat step 8 until all the mixture from **Step 7** has been transferred to the Mini Filter Spin column.
  10. Apply 400  $\mu$ L of **Wash Solution** to the column and centrifuge for **1 minute at 14,000 RPM**. Discard the flowthrough and reassemble the spin column with its collection tube.
  11. Repeat step 10 two more times, for a total of three washes.
  12. Spin the column, empty, for **3 minutes at 14,000 RPM**. Discard the collection tube.
  13. Transfer the spin column to a fresh 1.7 mL Elution tube. Apply 100  $\mu$ L of **Elution solution** to the column and centrifuge for **2 minutes at 2,000 RPM**, followed by **3 minute at 14,000 RPM**.
- ❖ **Free-circulating plasma/serum RNA is now ready for downstream applications.**

**Table 1. PS Solution A, PS Solution B and 96-100% Ethanol to be added to different Plasma/Serum sample volumes**

| Sample Volume (mL) | PS Solution A (mL)<br>(Step 1) | PS Solution B (mL)<br>(Step 1) | 96-100% Ethanol (mL)<br>(Step 3) |
|--------------------|--------------------------------|--------------------------------|----------------------------------|
| 2                  | 0.2                            | 3.8                            | 6                                |
| 3                  | 0.2                            | 5.8                            | 9                                |
| 4                  | 0.2                            | 7.8                            | 12                               |
| 5                  | 0.2                            | 9.8                            | 15                               |

## **Frequently Asked Questions**

### **1. What If a variable speed centrifuge is not available?**

- A fixed speed centrifuge can be used, however reduced yields may be observed.

### **2. What will happen if my centrifugation speed varied from the recommended speed?**

- This may lead to the degradation of the isolated RNA or reduction in the total RNA yields.

### **3. At what temperature should I centrifuge my samples?**

- All centrifugation steps are performed at room temperature. Centrifugation at 4°C will not adversely affect kit performance.

### **4. Can I process a different Plasma/Serum volume?**

- Yes, you can. To process different Plasma/Serum volumes please check Table 1. for the appropriate volumes that should be added from **PS Solution A**, **PS Solution B** and **96-100% Ethanol** to different Plasma/Serum sample volumes. The volume of **PS Solution C** is fixed for all Plasma/Serum volumes.

### **5. What If I added more or less of the specified reagents' volume?**

- Adding more or less from the specified volumes outlined in Table 1 may affect both the quality and quantity of the isolated RNA.

### **6. What If I forgot to do a dry spin after my second wash?**

- Your elution will be contaminated with the Wash Solution that contains Ethanol. This will dilute the RNA yield and it will interfere with your downstream applications.

### **7. Can I perform a second elution?**

- Yes, you can. A second elution is possible, but it is recommended that this elution is performed in a smaller volume (50 µL).

### **8. Why do my samples show low RNA yield?**

- Plasma/Serum samples contain very little RNA. This varies from individual to individual based on numerous variables. In order to increase the yield, the amount of Plasma/Serum input could be increased.

### **9. Why do the A260:280 ratio of the purified RNA is lower than 2.0?**

- Most of the Free-Circulating Plasma/Serum RNA is degraded and present in short fragment. The A260:280 ratio is normally between 1 – 1.6. This low A260:280 ratio will not affect any downstream application

### **10. Why my isolated RNA do not perform well in downstream applications?**

- If a different Elution Buffer was used other than the one provided in the kit, the buffer should be checked for any components that may interfere with the application. Common components that are known to interfere are high salts (including EDTA), detergents and other denaturants. Check the compatibility of your elution buffer with the intended use.

### **11. Do I need to do a DNase treatment for my RNA Elution?**

- You may need to do a DNase treatment to your isolated Plasma/Serum Circulating RNA. It is recommended to use Norgen's RNase-Free DNase I Kit (Cat# 25710)

| <b>Related Products</b>                    | <b>Product #</b> |
|--|------------------|
| Plasma/Serum Circulating RNA Isolation Kit | 30000            |
| RNase-Free DNase I Kit                     | 25710            |
| Total RNA Purification Kit                 | 17200            |
| Total RNA Purification 96-well Kit         | 24300            |
| Blood Genomic DNA Isolation Kit            | 18200            |

**Technical Assistance**

NORGEN's Technical Service Department is staffed by experienced scientists with extensive practical and theoretical expertise in sample and assay technologies and the use of NORGEN products. If you have any questions or experience any difficulties regarding Norgen's Plasma/Serum Circulating RNA Purification Kit (Slurry Format) or NORGEN products in general, please do not hesitate to contact us.

NORGEN customers are a valuable source of information regarding advanced or specialized uses of our products. This information is helpful to other scientists as well as to the researchers at NORGEN. We therefore encourage you to contact us if you have any suggestions about product performance or new applications and techniques.

For technical assistance and more information, please contact our Technical Support Team between the hours of 8:30 and 5:30 (Eastern Standard Time) at (905) 227-8848 or Toll Free at 1-866-667-4362. or call one of the NORGEN local distributors ([www.norgenbiotek.com](http://www.norgenbiotek.com)) or through email at [techsupport@norgenbiotek.com](mailto:techsupport@norgenbiotek.com).

3430 Schmon Parkway, Thorold, ON Canada L2V 4Y6  
Phone: (905) 227-8848  
Fax: (905) 227-1061  
Toll Free in North America: 1-866-667-4362